



Manaaki Whenua
Landcare Research

Gerald Street
P.O.Box 69040
Lincoln, 7640
Ph: +64 3 321 9999
Fax: +64 3 321 9998

Toxicology Laboratory Analysis Report

Report No: T5993

CLIENT: [REDACTED], Watercare Services Ltd, Private Bag 92-521, Wellesley Street Auckland 1141

CLIENT REFERENCE No.: PO 3000207709

Telephone No: 09 539 7300

SAMPLES: One water sample

REQUIREMENT: Examine for 1080

RECEIVED: 16 September 2015

Sample/s were received for analysis. The details were entered into the laboratory sample system and the sample/s given a reference number. The sample details and results are as follows:

No. samples: 1

LabNo.	Description	1080, µg/mL
19724	Water sample, 150915-156-1, Mangatangi Valve Tower 1300, 15/9/15, 1406	<MDL

All results are reported to two significant figures.

The determination was carried out using TLM005, the assay of 1080 in water, soil and biological materials by GLC. The method detection limit (MDL) is 0.0001 µg/mL and the uncertainty (95% c.i.) is ± 12%.

TESTED BY: leb **WORKBOOK REF:** 98/9

TEST PERIOD: 16-17/9/15

AUTHORISED BY:

L.H.Booth, L.E. Brown

Date: 17/09/2015



These results are confidential to the client and relate only to the samples as received and tested. This report may be reproduced in full only. The samples relating to this report will be disposed of after two months from the report date unless requested otherwise by the client. Where appropriate, the above results will be included in anonymised form in the National Vertebrate Pesticide Residue Database.